GROUPE APOCOM

MIT MEDICAL INTERNATIONAL TECHNOLOGIES (MIT CANADA) INC.

ANNUAL SERVICE AND MAINTENANCE

MESO-JET MULTIJET

INSPECTION & FUNCTIONALITY TEST

1. Inspect how the system was packaged for shipping (photo taken).
2. Inspect the entire system outside for cleanliness and any damage (photo taken).
3. Injection test to analyze leaks of air or saline from the system.
4. Observe the injection power coming out of the nozzle holder.
5. Inspect the operation of the multijet barrel in comparison to the dimming and piston of the disposable cartridge.

GENERAL MAINTENANCE

1. Cleaning the injector, including the multijet barrel.
2. Changes the bullet for click sound, if necessary.
3. Verification of calibration.



COMPONENT COST = $250.00
LABOUR COST = 3 hours x $125.00 = $375.00
TOTAL COST = $625.00*

*Shipping costs not included

REMARKS :

- If, after the inspection, major components not specified above need to be replaced, Groupe Apocom will provide the customer with the cost of the service, including components and labor, and will request authorization before proceeding with the service.
- Please note that shipping costs to the MIT Canada offices and for the return are the responsibility of the customer. You can take care of the transport yourself or benefit from the advantageous rates of Groupe Apocom's carriers by making a request.

Package size : ____ inches wide X ____ inches length X ____ inches high **Package weight :** ____ lbs

Here are the parts you need to send in your package for maintenance.



Customer name : _____ Tel. : _____ I agree : ☐

Cie name : _____ Date : _____ Client's signature : _____

www.groupeapocom.com

Send by email: info@groupeapocom.com or fax: 450 438-8113

1 800 813-3338 | 514 259-5901 | info@groupeapocom.com